
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001600
11:02 11/19/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:40
PRESSURIZING 6:23
STERILIZING 7:56
136.0 C 220kPa 7:56
136.0 C 220kPa 8:26
136.4 C 218kPa 8:56
136.5 C 222kPa 9:26
136.4 C 222kPa 9:56
135.9 C 220kPa 10:26
136.5 C 223kPa 10:56
136.2 C 221kPa 11:26
VENTING 11:26
AIR DRYING 12:26
CYCLE COMPLETE 42:26

Digital Signature #
6931D5D192974E47
