

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4145

Name: RITTER 1  
Sterilizer ID: V2178198  
12/21/2023 08:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.7 F 0.0  
2:00 74.2 F 0.3  
4:00 92.3 F 1.3  
6:00 127.6 F 3.7  
8:00 153.1 F 6.6  
10:00 176.8 F 10.2  
12:00 195.3 F 8.1  
14:00 210.8 F 4.1  
16:00 220.2 F 7.1  
18:00 229.7 F 10.6  
20:00 233.1 F 8.2  
22:00 239.0 F 10.1  
24:00 238.0 F 8.9  
26:00 246.1 F 12.8  
28:00 257.0 F 18.5  
30:00 266.6 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 27.8  
1:00 272.4 F 28.6  
2:00 271.1 F 27.7  
3:00 272.4 F 28.5  
4:00 271.1 F 27.7

Min 271.0 F 27.7  
Max 272.6 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 249.7 F 0.0  
10:00 225.9 F 0.0  
15:00 224.3 F 0.0  
20:00 226.7 F 0.0  
25:00 233.1 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:30  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:09:43  
CYCLE COMPLETE

12/21/2023 10:06 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_