

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007439  
6:35 28/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:54  
STERILIZING 4:05  
136.4 C 222kPa 4:05  
135.8 C 219kPa 4:35  
136.1 C 221kPa 5:05  
136.1 C 221kPa 5:35  
136.4 C 217kPa 6:05  
136.4 C 221kPa 6:35  
136.2 C 221kPa 7:05  
136.3 C 222kPa 7:35  
136.3 C 222kPa 7:35  
VENTING 7:36  
AIR DRYING 8:28  
CYCLE COMPLETE 38:28

-----  
Digital Signature #  
1E41695FF271EAD0  
-----