
Midmark M11 - v1.0.0
Total Cycles: 3463

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2023 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.3 F 0.0
2:00 173.2 F 0.6
4:00 170.1 F 3.6
6:00 197.2 F 7.6
8:00 218.6 F 13.0
10:00 225.9 F 7.2
12:00 237.0 F 12.2
14:00 234.9 F 8.9
16:00 241.4 F 10.2
18:00 246.0 F 13.2
20:00 258.0 F 19.6
22:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.2
2:00 272.4 F 28.8
3:00 271.5 F 28.2
4:00 272.0 F 28.9

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 232.2 F 0.0
10:00 224.8 F 0.0
15:00 224.8 F 0.0
20:00 232.1 F 0.0
25:00 223.8 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:20
CYCLE COMPLETE

06/29/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____