

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000180  
8:08 17/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:30  
PRESSURIZING 4:30  
STERILIZING 5:42  
136.0 C 220kPa 5:42  
136.4 C 225kPa 6:12  
136.3 C 224kPa 6:42  
136.3 C 225kPa 7:12  
136.4 C 226kPa 7:42  
136.4 C 226kPa 8:12  
136.4 C 226kPa 8:42  
136.4 C 226kPa 9:12  
136.4 C 226kPa 9:12  
VENTING 9:12  
AIR DRYING 10:15  
CYCLE COMPLETE 25:15  
-----

Digital Signature #  
FF2136FB2F2C6CF2  
-----