
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001408
11:59 03/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:17
STERILIZING 3:15
136.0 C 220kPa 3:15
136.1 C 222kPa 3:45
135.9 C 221kPa 4:15
136.4 C 225kPa 4:45
136.1 C 221kPa 5:15
136.4 C 223kPa 5:45
136.4 C 224kPa 6:15
136.4 C 224kPa 6:45
136.4 C 224kPa 6:45
VENTING 6:45
AIR DRYING 7:34
CYCLE COMPLETE 22:34

Digital Signature #
601E5B28EEAC3DFA
