
Midmark M11 - v1.0.5
Total Cycles: 3027

Name: RITTER2
Sterilizer ID: V2313228
10/18/2023 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.2 F 0.0
2:00 139.0 F 1.8
4:00 159.4 F 5.0
6:00 183.6 F 9.4
8:00 219.1 F 7.2
10:00 227.9 F 8.9
12:00 237.4 F 10.8
14:00 239.6 F 11.1
16:00 236.5 F 9.3
18:00 250.7 F 15.8
20:00 262.3 F 22.6
22:00 272.8 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.9 F 29.3
2:00 272.4 F 29.6
3:00 272.6 F 29.9
4:00 271.4 F 29.3

Min 271.4 F 29.2
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 251.1 F 0.0
10:00 194.5 F 0.0
15:00 187.5 F 0.0
20:00 184.1 F 0.0
25:00 181.7 F 0.0
30:00 180.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:01:19
CYCLE COMPLETE

10/18/2023 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____