
Midmark M11 - v1.0.5
Total Cycles: 1741

Name: RITTER2
Sterilizer ID: V2313228
11/17/2022 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 181.7 F 1.1
4:00 192.3 F 5.8
6:00 214.8 F 11.5
8:00 224.5 F 7.2
10:00 238.8 F 13.0
12:00 240.4 F 11.6
14:00 240.8 F 11.1
16:00 255.1 F 18.4
18:00 267.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.4
2:00 271.6 F 29.0
3:00 273.0 F 29.6
4:00 272.0 F 28.9

Min 271.5 F 28.7
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.9 F 0.0
5:00 210.3 F 0.0
10:00 227.0 F 0.0
15:00 227.5 F 0.0
20:00 229.2 F 0.0
25:00 227.8 F 0.0
30:00 225.5 F 0.0

Drying Complete

FILLING: 1:20
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:51
CYCLE COMPLETE

11/17/2022 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____