

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2011

Name: RITTER 1  
Sterilizer ID: V2178198  
06/08/2022 02:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
2:00 177.1 F 4.5  
4:00 191.0 F 9.1  
6:00 206.1 F 8.5  
8:00 221.5 F 7.4  
10:00 234.0 F 12.4  
12:00 235.6 F 9.5  
14:00 237.0 F 8.9  
16:00 247.9 F 14.4  
18:00 260.1 F 21.2  
20:00 271.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.4  
2:00 272.4 F 29.0  
3:00 271.5 F 28.4  
4:00 271.6 F 28.9  
  
Min 271.2 F 28.4  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.3 F 0.0  
5:00 223.1 F 0.0  
10:00 230.7 F 0.0  
15:00 232.9 F 0.0  
20:00 230.3 F 0.0  
25:00 219.3 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 20:47  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

06/08/2022 03:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_