

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1527

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2022 02:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.9 F 0.0  
2:00 207.3 F 4.2  
4:00 220.9 F 9.7  
6:00 230.3 F 7.9  
8:00 244.1 F 12.9  
10:00 241.2 F 11.4  
12:00 239.4 F 10.3  
14:00 253.4 F 17.4  
16:00 265.3 F 24.5  
18:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.4 F 29.4  
1:00 271.4 F 29.3  
2:00 271.2 F 29.6  
3:00 271.3 F 29.7  
4:00 271.2 F 29.7  
  
Min 271.0 F 29.3  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.4 F 0.0  
5:00 237.4 F 0.0  
10:00 228.8 F 0.0  
15:00 229.8 F 0.0  
20:00 231.5 F 0.0  
25:00 231.1 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 2:30  
HEATING: 18:13  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

09/22/2022 03:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_