
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007405
12:05 21/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:42
STERILIZING 4:46
136.0 C 220kPa 4:46
136.4 C 225kPa 5:16
136.1 C 223kPa 5:46
136.4 C 225kPa 6:16
136.2 C 221kPa 6:46
136.4 C 223kPa 7:16
136.4 C 223kPa 7:46
136.4 C 224kPa 8:16
136.4 C 224kPa 8:16
VENTING 8:16
AIR DRYING 9:15
CYCLE COMPLETE 34:17

Digital Signature #
DE91F1BA2D193BB6
