
Midmark M11 - v1.0.5
Total Cycles: 1356

Name: RITTER2
Sterilizer ID: V2313228
07/27/2022 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 186.2 F 0.3
4:00 198.0 F 3.5
6:00 216.5 F 8.2
8:00 233.7 F 13.4
10:00 235.9 F 9.9
12:00 237.9 F 9.0
14:00 243.3 F 12.2
16:00 245.8 F 13.3
18:00 258.2 F 20.0
20:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.3 F 29.3
2:00 271.4 F 28.5
3:00 272.2 F 29.4
4:00 271.9 F 28.9

Min 271.1 F 28.5
Max 272.3 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 233.0 F 0.0
10:00 228.9 F 0.0
15:00 228.5 F 0.0
20:00 226.8 F 0.0
25:00 228.1 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

07/27/2022 01:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____