
Midmark M11 - v1.0.5
Total Cycles: 1697

Name: RITTER2
Sterilizer ID: V2313228
11/08/2022 10:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 179.3 F 1.3
4:00 196.4 F 6.0
6:00 214.1 F 11.4
8:00 226.8 F 7.5
10:00 240.7 F 13.4
12:00 241.3 F 11.7
14:00 240.8 F 10.8
16:00 254.9 F 18.0
18:00 266.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.1 F 28.8
2:00 272.6 F 29.3
3:00 271.6 F 28.5
4:00 272.4 F 29.1

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 232.2 F 0.0
10:00 225.9 F 0.0
15:00 223.8 F 0.0
20:00 227.7 F 0.0
25:00 228.9 F 0.0
30:00 226.1 F 0.0

Drying Complete

FILLING: 1:36
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:09
CYCLE COMPLETE

11/08/2022 11:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____