
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008248
15:49 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 4:02
136.5 C 222kPa 4:02
136.1 C 223kPa 4:32
136.2 C 224kPa 5:02
136.4 C 223kPa 5:32
136.4 C 223kPa 6:02
136.4 C 224kPa 6:32
136.2 C 222kPa 7:02
136.3 C 222kPa 7:32
136.3 C 222kPa 7:32
VENTING 7:33
AIR DRYING 8:27
CYCLE COMPLETE 8:49

Digital Signature #
EA0ED67FA0FD05C0
