
Midmark M11 - v1.0.0
Total Cycles: 3317

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2023 07:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 80.2 F 0.0
4:00 86.8 F 0.0
6:00 112.2 F 0.0
8:00 140.5 F 0.2
10:00 172.7 F 3.3
12:00 196.3 F 6.6
14:00 205.1 F 10.4
16:00 217.1 F 7.0
18:00 224.5 F 7.7
20:00 235.8 F 11.9
22:00 233.7 F 8.3
24:00 243.8 F 12.1
26:00 244.1 F 12.2
28:00 255.7 F 18.3
30:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 29.0
2:00 271.9 F 28.3
3:00 271.7 F 28.4
4:00 272.7 F 28.9

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 203.3 F 0.0
10:00 202.6 F 0.0
15:00 198.0 F 0.0
20:00 197.2 F 0.0
25:00 209.2 F 0.0
30:00 207.8 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

06/01/2023 08:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____