
Midmark M11 - v1.0.5
Total Cycles: 918

Name: RITTER2
Sterilizer ID: V2313228
03/29/2022 07:06 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.5 F 0.0
2:00 75.8 F 0.0
4:00 82.3 F 0.0
6:00 117.7 F 0.6
8:00 144.9 F 3.5
10:00 171.8 F 6.7
12:00 194.2 F 10.3
14:00 213.9 F 7.1
16:00 222.5 F 8.6
18:00 235.7 F 13.1
20:00 235.9 F 10.0
22:00 237.3 F 8.9
24:00 247.4 F 14.4
26:00 258.8 F 20.8
28:00 269.4 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.6
1:00 272.3 F 29.6
2:00 272.6 F 29.9
3:00 272.5 F 30.2
4:00 271.7 F 29.8

Min 271.5 F 29.5
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 221.4 F 0.0
10:00 211.5 F 0.0
15:00 214.9 F 0.0
20:00 212.6 F 0.0
25:00 214.2 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:47
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:07:41
CYCLE COMPLETE

03/29/2022 08:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____