
Midmark M11 - v1.0.0
Total Cycles: 2042

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.9 F 0.0
2:00 76.9 F 0.0
4:00 86.2 F 0.0
6:00 116.5 F 0.6
8:00 149.5 F 2.2
10:00 174.3 F 4.6
12:00 191.9 F 7.7
14:00 207.2 F 11.0
16:00 222.9 F 5.4
18:00 228.4 F 7.7
20:00 237.2 F 11.7
22:00 233.7 F 8.0
24:00 244.9 F 12.5
26:00 244.0 F 12.1
28:00 255.9 F 18.3
30:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.9
2:00 271.7 F 28.1
3:00 272.1 F 28.8
4:00 272.4 F 28.6

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 219.8 F 0.0
10:00 197.2 F 0.0
15:00 226.7 F 0.0
20:00 246.2 F 0.0
25:00 228.8 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:33
CYCLE COMPLETE

06/22/2022 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____