

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5037

Name: RITTER 1  
Sterilizer ID: V2178198  
10/30/2024 10:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 115.7 F 0.0  
2:00 113.9 F 0.0  
4:00 130.3 F 0.4  
6:00 157.7 F 0.8  
8:00 181.2 F 3.7  
10:00 196.3 F 7.3  
12:00 212.5 F 10.9  
14:00 223.0 F 5.8  
16:00 228.3 F 8.7  
18:00 239.4 F 13.4  
20:00 239.5 F 10.7  
22:00 235.8 F 8.6  
24:00 249.6 F 14.9  
26:00 260.5 F 21.1  
28:00 270.7 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.3 F 28.9  
2:00 272.0 F 28.5  
3:00 271.2 F 28.0  
4:00 271.8 F 28.8  
  
Min 271.1 F 28.0  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 193.0 F 0.0  
10:00 188.4 F 0.0  
15:00 199.4 F 0.0  
20:00 207.5 F 0.0  
25:00 213.9 F 0.0  
30:00 213.2 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 28:52  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:07:26  
CYCLE COMPLETE

10/30/2024 11:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_