
Midmark M11 - v1.0.0
Total Cycles: 3203

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 174.5 F 2.6
4:00 191.5 F 6.7
6:00 212.9 F 11.9
8:00 222.2 F 7.1
10:00 236.8 F 12.5
12:00 236.8 F 9.9
14:00 238.1 F 8.4
16:00 250.0 F 15.2
18:00 262.7 F 22.5
20:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 29.0
2:00 271.4 F 28.2
3:00 272.5 F 29.2
4:00 272.1 F 28.7

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 208.6 F 0.0
10:00 225.3 F 0.0
15:00 228.6 F 0.0
20:00 227.5 F 0.0
25:00 224.7 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 20:09
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

05/04/2023 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____