

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1673

Name: RITTER2  
Sterilizer ID: V2313228  
11/02/2022 05:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.0 F 0.0  
2:00 165.8 F 2.9  
4:00 175.1 F 6.9  
6:00 206.5 F 12.0  
8:00 221.4 F 7.0  
10:00 232.3 F 12.6  
12:00 236.1 F 9.9  
14:00 237.6 F 8.8  
16:00 249.4 F 15.3  
18:00 261.4 F 22.2  
20:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.5  
2:00 271.5 F 29.0  
3:00 272.6 F 29.6  
4:00 271.7 F 28.9

Min 271.4 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 220.6 F 0.0  
10:00 238.4 F 0.0  
15:00 244.8 F 0.0  
20:00 233.5 F 0.0  
25:00 223.1 F 0.0  
30:00 218.7 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

11/02/2022 06:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_