
Midmark M11 - v1.0.0
Total Cycles: 3544

Name: RITTER 1
Sterilizer ID: V2178198
07/18/2023 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 184.2 F 1.3
4:00 191.8 F 4.9
6:00 212.0 F 10.0
8:00 222.4 F 6.3
10:00 233.5 F 10.6
12:00 233.0 F 7.5
14:00 244.0 F 12.3
16:00 245.1 F 12.8
18:00 258.2 F 19.8
20:00 269.6 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.8
2:00 271.7 F 28.6
3:00 272.5 F 28.8
4:00 271.6 F 28.4

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 209.7 F 0.0
10:00 214.8 F 0.0
15:00 219.7 F 0.0
20:00 221.2 F 0.0
25:00 221.2 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

07/18/2023 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____