
Midmark M11 - v1.0.5
Total Cycles: 3295

Name: RITTER2
Sterilizer ID: V2313228
01/12/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.2 F 0.0
2:00 103.0 F 0.2
4:00 119.8 F 1.6
6:00 149.8 F 4.7
8:00 177.6 F 8.1
10:00 196.2 F 11.9
12:00 216.1 F 5.1
14:00 227.1 F 9.4
16:00 233.3 F 10.5
18:00 238.5 F 11.1
20:00 236.7 F 9.3
22:00 249.5 F 15.5
24:00 260.7 F 21.8
26:00 270.8 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.3 F 30.0
2:00 271.6 F 29.6
3:00 271.9 F 29.4
4:00 272.7 F 29.7

Min 270.7 F 28.8
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 198.7 F 0.0
10:00 189.0 F 0.0
15:00 188.8 F 0.0
20:00 187.2 F 0.0
25:00 185.9 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:43
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:05:17
CYCLE COMPLETE

01/12/2024 04:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____