

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006996  
12:37 27/02/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:17  
PRESSURIZING 3:18  
STERILIZING 4:20  
136.0 C 220kPa 4:20  
136.4 C 224kPa 4:50  
136.1 C 218kPa 5:20  
136.4 C 224kPa 5:50  
136.4 C 225kPa 6:20  
135.8 C 220kPa 6:50  
136.4 C 223kPa 7:20  
136.4 C 223kPa 7:50  
136.4 C 223kPa 7:50  
VENTING 7:50  
AIR DRYING 8:45  
CYCLE COMPLETE 8:48

-----  
Digital Signature #  
758243BE2913483D  
-----