
Midmark M11 - v1.0.0
Total Cycles: 3757

Name: RITTER 1
Sterilizer ID: V2178198
09/12/2023 10:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.5 F 0.0
2:00 171.8 F 0.2
4:00 185.3 F 3.1
6:00 215.9 F 7.4
8:00 219.4 F 12.3
10:00 232.5 F 7.9
12:00 242.4 F 13.2
14:00 240.7 F 10.9
16:00 237.7 F 9.3
18:00 251.7 F 15.9
20:00 263.1 F 22.6
22:00 272.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.9 F 29.1
2:00 271.8 F 28.4
3:00 272.7 F 29.1
4:00 271.7 F 28.4

Min 270.9 F 27.9
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 228.6 F 0.0
10:00 226.5 F 0.0
15:00 211.6 F 0.0
20:00 218.8 F 0.0
25:00 218.1 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:14
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:38
CYCLE COMPLETE

09/12/2023 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____