
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002454
10:22 27/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:16
136.0 C 220kPa 4:16
136.3 C 224kPa 4:46
136.4 C 225kPa 5:16
136.4 C 225kPa 5:46
136.3 C 223kPa 6:16
136.4 C 223kPa 6:46
136.1 C 222kPa 7:16
136.4 C 225kPa 7:46
136.4 C 225kPa 7:46
VENTING 7:46
AIR DRYING 8:39
CYCLE COMPLETE 8:50

Digital Signature #
279943E46711BAA6
