
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 000295
9:12 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:32
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:32

Digital Signature #
98CDEF79CC4AD151
