
Midmark M11 - v1.0.0
Total Cycles: 1776

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.7 F 0.0
2:00 202.6 F 3.8
4:00 212.9 F 9.0
6:00 223.3 F 7.5
8:00 235.5 F 10.7
10:00 235.2 F 8.2
12:00 240.5 F 10.3
14:00 247.1 F 14.1
16:00 260.0 F 21.1
18:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.5
2:00 272.2 F 28.9
3:00 271.3 F 28.7
4:00 272.5 F 29.6

Min 271.2 F 28.5
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 226.1 F 0.0
10:00 229.9 F 0.0
15:00 236.4 F 0.0
20:00 232.6 F 0.0
25:00 230.7 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

03/31/2022 11:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____