
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001544
10:27 10/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:47
PRESSURIZING 6:29
STERILIZING 8:02
136.0 C 220kPa 8:02
135.8 C 218kPa 8:32
136.4 C 222kPa 9:02
136.4 C 222kPa 9:33
136.2 C 220kPa 10:03
135.7 C 219kPa 10:33
136.2 C 221kPa 11:03
136.4 C 219kPa 11:32
VENTING 11:32
AIR DRYING 12:36
CYCLE COMPLETE 42:36

Digital Signature #
FA07F894F6DC93FA
