
Midmark M11 - v1.0.5
Total Cycles: 1806

Name: RITTER2
Sterilizer ID: V2313228
12/05/2022 06:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.2 F 0.0
2:00 152.8 F 2.7
4:00 172.7 F 6.4
6:00 194.9 F 11.5
8:00 215.6 F 6.1
10:00 230.0 F 11.5
12:00 232.7 F 8.9
14:00 238.3 F 9.9
16:00 247.3 F 14.5
18:00 260.0 F 21.6
20:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.5
2:00 272.3 F 30.0
3:00 271.8 F 29.2
4:00 272.5 F 30.0

Min 271.5 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 234.0 F 0.0
10:00 231.2 F 0.0
15:00 230.1 F 0.0
20:00 229.1 F 0.0
25:00 223.0 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 20:34
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

12/05/2022 07:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____