
Midmark M11 - v1.0.5
Total Cycles: 2344

Name: RITTER2
Sterilizer ID: V2313228
05/05/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.5 F 0.0
2:00 175.0 F 2.5
4:00 183.4 F 7.3
6:00 203.8 F 13.3
8:00 226.0 F 8.4
10:00 236.3 F 9.1
12:00 243.7 F 11.8
14:00 247.3 F 13.9
16:00 261.2 F 21.7
18:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.3 F 28.7
2:00 272.8 F 29.3
3:00 271.6 F 28.4
4:00 272.8 F 29.1

Min 271.5 F 28.3
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 223.7 F 0.0
Dry Cycle Aborted

FILLING: 1:17
HEATING: 18:17
STERILIZING: 4:00
VENTING: 2:49
DRYING: 0:19
TOTAL CYCLE: 00:26:42

05/05/2023 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____