
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000822
17:00 13/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:49
STERILIZING 5:01
136.0 C 227kPa 5:01
136.3 C 231kPa 5:31
136.4 C 232kPa 6:01
136.4 C 232kPa 6:31
136.4 C 232kPa 7:01
136.3 C 232kPa 7:31
136.3 C 231kPa 8:01
136.4 C 230kPa 8:31
136.4 C 230kPa 8:31
VENTING 8:31
AIR DRYING 9:26
CYCLE COMPLETE 24:26

Digital Signature #
907B4CB093BBB1D9
