

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4483

Name: RITTER 1  
Sterilizer ID: V2178198  
04/24/2024 08:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.8 F 0.0  
2:00 81.2 F 0.2  
4:00 87.6 F 1.1  
6:00 114.3 F 3.2  
8:00 143.8 F 5.9  
10:00 168.7 F 9.2  
12:00 185.0 F 12.7  
14:00 203.9 F 3.8  
16:00 215.8 F 7.0  
18:00 225.2 F 10.7  
20:00 229.5 F 7.6  
22:00 237.2 F 10.6  
24:00 234.9 F 8.0  
26:00 247.0 F 13.9  
28:00 258.2 F 19.9  
30:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.4 F 29.2  
2:00 271.5 F 28.3  
3:00 272.5 F 29.3  
4:00 271.9 F 28.7

Min 271.3 F 28.3  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 219.2 F 0.0  
10:00 224.6 F 0.0  
15:00 227.7 F 0.0  
20:00 218.6 F 0.0  
25:00 218.9 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:20  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:35  
CYCLE COMPLETE

04/24/2024 09:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_