
Midmark M11 - v1.0.5
Total Cycles: 1980

Name: RITTER2
Sterilizer ID: V2313228
01/30/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 190.4 F 4.2
4:00 204.9 F 9.0
6:00 214.2 F 7.7
8:00 230.6 F 10.1
10:00 231.5 F 7.9
12:00 241.4 F 12.0
14:00 244.9 F 13.3
16:00 258.1 F 20.4
18:00 269.5 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.5 F 29.7
2:00 271.6 F 29.5
3:00 272.5 F 29.8
4:00 271.5 F 29.2

Min 271.4 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 221.6 F 0.0
10:00 220.3 F 0.0
15:00 232.1 F 0.0
20:00 227.3 F 0.0
25:00 225.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:56
CYCLE COMPLETE

01/30/2023 03:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____