
Midmark M11 - v1.0.5
Total Cycles: 3558

Name: RITTER2
Sterilizer ID: V2313228
04/25/2024 08:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.0 F 0.0
2:00 75.2 F 0.0
4:00 83.8 F 0.5
6:00 109.9 F 2.6
8:00 140.8 F 5.2
10:00 169.0 F 8.1
12:00 191.0 F 11.1
14:00 210.1 F 4.2
16:00 217.0 F 7.5
18:00 227.2 F 11.3
20:00 226.9 F 6.5
22:00 239.0 F 11.9
24:00 236.9 F 9.5
26:00 249.5 F 15.4
28:00 259.8 F 21.4
30:00 269.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.8 F 29.6
2:00 272.6 F 29.6
3:00 271.4 F 29.0
4:00 272.3 F 30.0

Min 271.2 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 215.5 F 0.0
10:00 211.9 F 0.0
15:00 215.4 F 0.0
20:00 215.9 F 0.0
25:00 226.3 F 0.0
30:00 234.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:55
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:01
CYCLE COMPLETE

04/25/2024 09:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____