
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007305
12:01 02/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:55
STERILIZING 2:53
136.0 C 220kPa 2:53
136.4 C 222kPa 3:23
136.4 C 223kPa 3:53
136.4 C 223kPa 4:23
136.4 C 224kPa 4:53
136.4 C 224kPa 5:23
136.4 C 222kPa 5:53
136.4 C 223kPa 6:23
136.4 C 223kPa 6:23
VENTING 6:23
AIR DRYING 7:15
CYCLE COMPLETE 37:15

Digital Signature #
7186D37448D696C2
