
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000778
17:02 16/02/2022
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 4:06
CYCLE ABORTED 12:31
WATER LEVEL LOW
NOT STERILE
END TIME 12:31

Digital Signature #
BD2DE025B3768972
