

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000485  
10:49 26/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:20  
STERILIZING 4:34  
136.0 C 220kPa 4:34  
136.4 C 230kPa 5:04  
136.4 C 233kPa 5:34  
136.3 C 232kPa 6:04  
135.9 C 230kPa 6:34  
136.4 C 233kPa 7:04  
136.1 C 231kPa 7:34  
136.3 C 232kPa 8:04  
136.3 C 232kPa 8:04  
VENTING 8:04  
AIR DRYING 8:56  
CYCLE COMPLETE 11:31  
-----

Digital Signature #  
013CD03AEC6F652B  
-----