
Midmark M11 - v1.0.5
Total Cycles: 2014

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 04:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.0 F 0.0
2:00 182.2 F 3.3
4:00 201.8 F 8.1
6:00 220.1 F 12.5
8:00 230.2 F 9.3
10:00 235.6 F 9.3
12:00 243.3 F 12.2
14:00 246.2 F 13.6
16:00 259.4 F 20.9
18:00 271.1 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 271.6 F 29.2
4:00 272.7 F 29.5

Min 271.3 F 28.6
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 238.9 F 0.0
10:00 228.5 F 0.0
15:00 224.7 F 0.0
20:00 217.9 F 0.0
25:00 216.1 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 18:44
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

02/06/2023 05:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____