
Midmark M11 - v1.0.0
Total Cycles: 2444

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.8 F 0.0
2:00 173.1 F 3.5
4:00 189.3 F 7.9
6:00 206.2 F 13.1
8:00 219.0 F 7.4
10:00 234.0 F 12.4
12:00 234.9 F 9.6
14:00 237.2 F 8.9
16:00 247.6 F 14.4
18:00 259.3 F 21.0
20:00 270.5 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.1 F 29.3
2:00 271.3 F 29.1
3:00 272.1 F 29.3
4:00 271.1 F 28.9

Min 271.1 F 28.7
Max 272.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 209.2 F 0.0
10:00 217.6 F 0.0
15:00 221.2 F 0.0
20:00 229.7 F 0.0
25:00 234.2 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:50
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

10/17/2022 04:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____