

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1618

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2022 01:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
2:00 182.9 F 2.9  
4:00 200.8 F 7.9  
6:00 214.7 F 13.5  
8:00 226.6 F 9.6  
10:00 236.1 F 9.1  
12:00 242.6 F 12.5  
14:00 244.4 F 13.3  
16:00 257.3 F 20.3  
18:00 268.7 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 272.9 F 30.1  
2:00 272.0 F 29.4  
3:00 273.0 F 30.3  
4:00 272.7 F 29.9  
  
Min 271.8 F 29.4  
Max 273.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 204.6 F 0.0  
10:00 216.5 F 0.0  
15:00 224.3 F 0.0  
20:00 229.5 F 0.0  
25:00 230.4 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 3:58  
HEATING: 19:09  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:23  
CYCLE COMPLETE

10/18/2022 02:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_