
Midmark M11 - v1.0.5
Total Cycles: 3399

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 05:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.8 F 0.0
2:00 137.3 F 0.6
4:00 151.2 F 3.3
6:00 174.4 F 7.3
8:00 196.3 F 11.6
10:00 215.2 F 5.2
12:00 225.2 F 9.7
14:00 232.1 F 9.5
16:00 240.0 F 12.1
18:00 239.5 F 10.8
20:00 252.6 F 17.3
22:00 263.9 F 23.9
24:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.8
2:00 271.4 F 29.1
3:00 271.6 F 29.8
4:00 272.1 F 29.6

Min 271.0 F 29.0
Max 272.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.3 F 0.0
5:00 210.3 F 0.0
10:00 224.5 F 0.0
15:00 230.7 F 0.0
20:00 229.0 F 0.0
25:00 229.1 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:03:02
CYCLE COMPLETE

02/21/2024 06:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____