

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005410  
8:35 24/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:07  
STERILIZING 4:12  
136.5 C 224kPa 4:12  
136.2 C 223kPa 4:42  
136.4 C 225kPa 5:12  
136.3 C 223kPa 5:42  
136.2 C 223kPa 6:12  
136.3 C 223kPa 6:42  
136.4 C 222kPa 7:12  
136.4 C 224kPa 7:42  
136.4 C 224kPa 7:42  
VENTING 7:42  
AIR DRYING 8:34  
CYCLE COMPLETE 11:25

-----  
Digital Signature #  
4BCDFF6840928C25  
-----