

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3187

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2023 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.7 F 0.0  
2:00 173.3 F 1.1  
4:00 191.9 F 4.8  
6:00 208.2 F 9.9  
8:00 223.5 F 7.1  
10:00 232.2 F 10.5  
12:00 232.0 F 7.8  
14:00 241.7 F 11.4  
16:00 245.4 F 13.2  
18:00 258.2 F 20.0  
20:00 269.5 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 271.7 F 29.0  
2:00 272.6 F 29.4  
3:00 271.9 F 28.9  
4:00 271.3 F 28.6  
  
Min 271.3 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
5:00 211.0 F 0.0  
10:00 232.2 F 0.0  
15:00 231.8 F 0.0  
20:00 223.2 F 0.0  
25:00 215.1 F 0.0  
30:00 211.1 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:56  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:59:30  
CYCLE COMPLETE

05/02/2023 12:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_