
Midmark M11 - v1.0.0
Total Cycles: 2049

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2022 07:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 75.5 F 0.0
4:00 86.4 F 0.0
6:00 114.2 F 0.6
8:00 145.5 F 2.2
10:00 173.5 F 4.4
12:00 189.7 F 7.2
14:00 203.2 F 10.6
16:00 216.9 F 6.7
18:00 223.7 F 7.3
20:00 233.6 F 11.4
22:00 231.4 F 7.3
24:00 243.9 F 12.9
26:00 241.6 F 11.1
28:00 253.8 F 17.2
30:00 264.4 F 23.5
32:00 271.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 29.0
2:00 271.5 F 28.1
3:00 272.4 F 29.0
4:00 272.1 F 28.6

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.7 F 0.0
5:00 208.9 F 0.0
10:00 226.3 F 0.0
15:00 231.0 F 0.0
20:00 221.8 F 0.0
25:00 216.7 F 0.0
30:00 212.7 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:05
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

06/24/2022 08:23 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____