
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002936
8:51 12/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:20
STERILIZING 4:27
136.5 C 223kPa 4:27
136.0 C 221kPa 4:57
136.3 C 221kPa 5:27
136.2 C 220kPa 5:57
136.4 C 224kPa 6:27
136.3 C 221kPa 6:57
136.2 C 223kPa 7:27
136.1 C 220kPa 7:57
136.1 C 220kPa 7:57
VENTING 7:57
AIR DRYING 8:49
CYCLE COMPLETE 8:54

Digital Signature #
3F95B40D583460CD
