
Midmark M11 - v1.0.0
Total Cycles: 1232

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2021 10:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.5 F 0.0
2:00 197.3 F 2.4
4:00 208.4 F 6.9
6:00 223.5 F 12.2
8:00 228.9 F 7.0
10:00 240.8 F 12.1
12:00 238.0 F 9.5
14:00 239.2 F 9.3
16:00 249.2 F 14.5
18:00 260.8 F 21.0
20:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 27.9
1:00 272.4 F 28.4
2:00 271.5 F 28.1
3:00 272.1 F 28.3
4:00 271.2 F 28.0

Min 271.2 F 27.8
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 234.6 F 0.0
10:00 226.6 F 0.0
15:00 206.8 F 0.0
20:00 210.9 F 0.0
25:00 213.7 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

10/05/2021 11:25 AM
ACCEPTED: X
REJECTED:
Operator: U01