

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3840

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2024 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.6 F 0.0  
2:00 141.6 F 1.8  
4:00 158.4 F 4.6  
6:00 179.1 F 8.5  
8:00 199.1 F 12.5  
10:00 219.4 F 5.0  
12:00 232.4 F 9.3  
14:00 238.4 F 11.5  
16:00 241.2 F 11.7  
18:00 239.5 F 10.4  
20:00 251.7 F 16.5  
22:00 262.3 F 22.7  
24:00 272.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.8 F 29.6  
2:00 271.8 F 29.6  
3:00 272.7 F 29.5  
4:00 271.8 F 29.6

Min 271.4 F 29.1  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 228.8 F 0.0  
10:00 225.7 F 0.0  
15:00 213.2 F 0.0  
20:00 206.7 F 0.0  
25:00 196.7 F 0.0  
30:00 196.5 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 25:19  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:06  
CYCLE COMPLETE

08/08/2024 12:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_