

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1207

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2021 11:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.7 F 0.0  
2:00 194.5 F 3.6  
4:00 218.5 F 8.6  
6:00 225.3 F 8.5  
8:00 233.5 F 9.8  
10:00 237.5 F 8.7  
12:00 244.9 F 13.5  
14:00 244.4 F 12.5  
16:00 257.1 F 19.2  
18:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.6 F 28.9  
2:00 271.7 F 28.5  
3:00 272.6 F 29.0  
4:00 271.6 F 28.3  
  
Min 271.5 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 218.7 F 0.0  
10:00 229.1 F 0.0  
15:00 224.9 F 0.0  
20:00 226.8 F 0.0  
25:00 230.2 F 0.0  
30:00 204.1 F 0.0  
  
Drying Complete

FILLING: 2:09  
HEATING: 19:12  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

09/27/2021 12:23 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01