
Midmark M11 - v1.0.0
Total Cycles: 2915

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 01:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.4 F 0.0
2:00 164.7 F 3.1
4:00 180.1 F 7.2
6:00 210.2 F 11.8
8:00 212.6 F 5.3
10:00 225.7 F 10.1
12:00 231.5 F 7.7
14:00 240.8 F 12.2
16:00 239.5 F 10.3
18:00 252.5 F 16.9
20:00 263.9 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.0 F 29.2
2:00 272.3 F 29.1
3:00 271.5 F 28.6
4:00 272.5 F 29.5

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 219.6 F 0.0
10:00 232.7 F 0.0
15:00 234.1 F 0.0
20:00 226.2 F 0.0
25:00 222.3 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

02/16/2023 02:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____