

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 002538  
7:24 27/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:26  
PRESSURIZING 1:30  
STERILIZING 2:35  
136.5 C 222kPa 2:35  
136.1 C 222kPa 3:05  
136.4 C 223kPa 3:35  
136.3 C 221kPa 4:05  
136.4 C 222kPa 4:35  
136.3 C 221kPa 5:05  
136.4 C 222kPa 5:35  
136.4 C 223kPa 6:05  
136.4 C 223kPa 6:05  
VENTING 6:05  
AIR DRYING 6:57  
CYCLE COMPLETE 7:15

-----  
Digital Signature #  
3A25764A26C13716  
-----